



- NOTES:**
1. ELECTRICAL REQUIREMENTS: 115 VAC, 60 Hz, SINGLE PHASE, 1Ø 20 AMP
 2. CONSTRUCTION:
WORKSURFACE/WORKAREA: 16GA. 304 SS, #4 FINISH
ACCESS PANELS, BLOWER SECTION: 16 & 18GA. C.R.S. BAKED ENAMEL FINISH, BAKER PERMA WHITE
 3. FILTER, MOTOR, BLOWER AND ALL ELECTRICAL COMPONENTS ARE FRONT ACCESSIBLE.
 4. PETCOCKS RECOMMENDED FOR USE WITH NONFLAMMABLE GAS, AIR, OR VACUUM
AT 30 P.S.I. MAXIMUM OPERATING PRESSURE.
SHOULD NATURAL GAS BE USED, PLEASE ADVISE OF ANY SPECIAL PLUMBING REQUIRED
BY YOUR LOCAL PLUMBING CODES; (ie BLACK IRON)
 5. CABINET PLUMBING CONNECTORS ARE 3/8 PARKER FITTINGS. 3/8 NPT ADAPTER
SHIPPED WITH CABINET.

EXHAUST REQUIREMENTS FOR SG303
WITH (HEC) HARD EXHAUST CONNECTION

SASH HEIGHT	EXHAUST FLOW RATE	PRESSURE LOSS
8"	162 CFM ±5	0.02" W.C.

NOTE: With optional Air tight damper add .10 to the above pressure loss.

OPTION #S	OPTION DESCRIPTION
19	(HEC) HARD EXHAUST CONNECTION
10	UV LIGHT OPTION
1	CHANNEL STAND



THE INFORMATION IN THIS DRAWING IS PROPRIETARY TO THE BAKER COMPANY INC. NO REPRODUCTION OR USE OF THE INFORMATION CONTAINED HEREIN MAY BE MADE WITHOUT EXPRESSED WRITTEN CONSENT OF THE BAKER COMPANY INC.

DO NOT SCALE DRAWING		THIS DRAWING AND ITS CONTENTS ARE THE PROPERTY OF THE BAKER COMPANY. THE INFORMATION CONTAINED HEREIN IS NOT TO BE REPRODUCED OR TRANSMITTED IN ANY FORM OR BY ANY MEANS, WITHOUT THE WRITTEN PERMISSION OF THE BAKER COMPANY.		THE BAKER COMPANY 885 GARDEN ST. BAKERSFIELD, CA 93311 TEL: (805) 833-1111 FAX: (805) 833-1112	
TITLE: STERILGUARD, MODEL SG303					
DESIGNED BY: AJohnson	DATE: 4/22/2003	SHEET NO: 1	MATERIAL: BGA-SG303-A2-1-10-19		REV: A
CHECKED BY: SCarpenter	DATE: 4/22/2003	SCALE: -	QUANTITY: -	1 OF 1	